

General & Useage information

Aqueous polymer based mountant for coverslipping glass microscope slides. Supplied in 100ml bottles.

RI approx 1.37 - 1.40.

Contains anti-fading agents and buffering reagents to maintain the pH at approx 8.5 – 8.80 .

Methods & results.

For coverslipping from water based reagents.

Make a hole in the top of the dispenser nozzle to allow a drop to form when the bottle is gently squeezed when upside down.

Onto the coverslip place 2-3 drops of reagent depending on the size of the coverslip being used. Place the coverslip on the subject matter on the slide and press down to ensure there are no bubbles under the coverslip. The slide can now be observed under oil immersion immediately or the slide can be left with the coverslip facing upwards to allow it to dry a little before storage. The mountant will dry slowly over several days in a warm environment (warm room temperature) to retain the coverslip if required to do so.

Storage:

Store reagent at cool room temperature in darkened cupboard away from direct sunlight. Shelf life 2 years.

No additional reagents required or quality controls.**Material safety Data Section**

Contains harmless plastic water soluble polymers, Tris buffer salts and 1,4-diazabicyclo[2,2,2]octane. If ingested seek immediate medical attention on what treatment is appropriate. Wash with soap and water in contact with skin. If in contact with eyes bathe eyes with eye-bath for 5 minutes. Seek medical attention if soreness persists. Wear suitable eye, face and body protection when using. This product is for "In-vitro diagnostic use only". Standard precautions for handling laboratory reagents should be followed when using this product.

S: 23-24-25

Waste disposal :

Clean the spillage area with detergent & water and run the waste to sewer drains.

Unused reagent can be disposed of by disposal to public sewer dilution greatly with tapwater.

Packaging should be rinsed with water and recycled as polyethylene, paper & cardboard.

Unsatisfactory performance

As part of our duty to monitor product performance and our policy of continual improvement. Please report to us any unsatisfactory performance you may experience with this product. If any reagent degrades before expiry of shelf life we will replace that reagent free of charge. GCC Diagnostics guarantees the quality of this product, the user should however determine the suitability of this product for their intended use.

GCC Diagnostics (Gainland Chemical Co) , Factory Rd, Sandycroft, Deeside, Flintshire. UK
Tel: 0044 1244 536326 Fax 00441244 531254 email gball@gccdiagnostics.com www.gccdiagnostics.com

