

General and useage information

Cytoplasmic fixative containing Mercuric chloride. This fixative has long been popular as a cytoplasmic fixative for wet smears.

It is not recommended for tissue as it can be hard in action and can cause excessive shrinkage to the tissue probably due to the excessive time needed for tissue penetration.

After fixation tissue or smears require treatment with iodine-alcohol and sodium thiosulphate to remove mercury deposits. Avaialble in packs of 500ml 1 litre .

Contains Mercuric chloride in an alcoholic base.

Storage:

Store reagent at cool room temperature in darkened cupboard . Shelf life 5 years.

No additional reagents or controls required.**Method**

Generally, wet smears are fixed for 10-20 minutes.

Smears or tissues fixed with any mercury containing fixative will require treatment to remove any brownish mercury precipitate which can be found in the specimen after fixation.

Treat with 0.25% iodine in 80% ethanol for 3 minutes

Rinse in tap water.

Place in 3% sodium thiosolphate (aq) for 3 minutes.

Wash in tap water for 1-2 minutes.

The Histopathology Laboratory will have their own in-house procedures for using well known fixatives such as Schaudinn's fixative.

Material safety data section

Contains Mercuric chloride which is TOXIC and CORROSIVE. Contains Ethanol which is Flammable. Do not consume this reagent. Use only in a well ventilled fume-hood. Do not breathe vapour. Use away from sources of ignition. If ingested seek immediate medical attention on what treatment is appropriate. Wash with soap and water in contact with skin. If in contact with eyes bathe eyes with eye-bath for 5 minutes. Seek medical attention if soreness persists. Wear suitable eye, face and body protection when using. This product is for "In-vitro diagnostic use only ". Standard precautions for handling laboratory reagents should be followed when usind this product.

R: 23-24-25-36-37-38 S: 23-24-25-36-37-38-51

Waste disposal

Clean the spillage area with detergent & water and remove waste to a licensed waste disposal contractor.

Unused reagent can be disposed of as shown above.

Packaging should be rinsed with water and recycled as polyethylene, paper & cardboard.

Unsatisfactory performance

As part of our duty to monitor product performance and our policy of continual improvement. Please report to us any unsatisfactory performance you may experience with this product. If any reagent degrades before expiry of shelf life we will replace that reagent free of charge. GCC Diagnostics guarantees the quality of this product, the user should however determine the suitability of this product for their intended use.

Gainland Chemical Co . GCC Diagnostics, Factory Rd, Sandycroft, Deeside, Flintshire. UK

Tel: 0044 1244 536326 Fax 00441244 531254 email gball@gccdiagnostics.com

www.gccdiagnostics.com

